

## Microvolume Protein Colorimetric Assay Data

### Technical Note 108

### Introduction

The optimal spectrophotometric method for determining the concentration of a protein is dependent upon both the protein and the sample buffers used. Direct measurement of the absorbance at 280 nm is generally applicable when working with purified protein solutions while colorimetric methods are routinely used for cell lysates and uncharacterized protein solutions.

Colorimetric assays are also used when the protein is suspended in a solution, such as a RIPA buffer, which has a strong absorbance signal in the UV range. Such buffers will interfere with the accuracy of a direct 280 nm measurement.

The DeNovix DS-11 spectrophotometer software includes a Colorimetrics app that is optimally configured for the 4 most common assays. The purpose of this technical note is present typical DS-11 data for each of the four assays included in the app.

### Materials

All assay reagents including phosphate buffered saline (PBS) (catalog #28372) and pre-diluted Bovine Serum Albumin (BSA) Standards (catalog #23208) were purchased from Fisher Scientific. The standards set included 0.125, 0.25, 0.50, 0.75, 1.0, 1.50 and 2.0 mg/ml solutions.

The colorimetric assay kits used in the studies were as follows:

- Pierce BCA Protein Assay Kit catalog #PI-23227
- Pierce Coomassie Plus (Bradford) Protein Assay Reagent catalog #PI-23238
- Pierce Modified Lowry Protein Assay Reagent Kit catalog #PI-23240
- Pierce 660nm Protein Assay Reagent catalog #PI-22660

### Assays

Each assay was performed according to the manufacturer's directions for the concentration range tested. In brief, tubes of the 7 pre-diluted protein standards along with a PBS buffer only sample were mixed with the respective assay reagents and incubated as described below.

#### BCA (Analysis at 562 nm)

The detergent compatible bicinchoninic acid assay (BCA assay) is used to quantify protein by measuring the Cu-BCA chelate formed in the presence of protein.

- Prepared working solution of 50:1 Reagent A:B. Added 1:20 sample volume to working solution. Incubated at 37°C for 30 minutes.

#### Bradford (Analysis at 595 nm)

The Bradford assay is based upon a protein-induced absorbance shift of a Coomassie Blue dye.

- Added 1:30 sample volume to assay reagent. Incubated solutions at RT for 10 minutes.

#### Lowry (Analysis at 750 nm)


The modified Lowry assay is used to quantify proteins based upon the absorbance of tetradentate copper-protein complexes.

- 1x Folin-Ciocalteu reagent was prepared in ultra pure dH<sub>2</sub>O. One mL of modified Lowry reagent was added to tubes with pre-diluted standards at 15 second intervals. Each tube was mixed well and incubated at room temperature for 10 minutes. 100µL of 1X Folin-Ciocalteu reagent was added to each tube and immediately vortexed. The tubes were covered and incubated at RT for 30 minutes.

#### Pierce 660 (Analysis at 660 nm)

The Pierce 660 nm Protein Assay measures a reddish-brown dye-metal complex that changes to green upon protein binding.

- Added 1:15 sample volume to assay reagent, then incubated at RT for 5 minutes.



## Microvolume Measurements

Upon launching the Colorimetrics app, the assay type of interest was selected before navigating to the Standard Curve screen. A Blank measurement was established using 1.5  $\mu\text{L}$  of the 0.00 mg/ml solution. A Standard Curve was then generated by measuring five 1.5  $\mu\text{L}$  aliquots of each standard concentration prepared as previously described.

Each measurement was done using a fresh aliquot of sample. The sample solution was removed between each measurements by wiping the upper and lower sample surfaces with a clean dry laboratory wipe.


## Results

An example curve is shown below for each of the four assays that are included in the Colorimetrics app. The data is presented as equivalent 10 mm absorbance values.

Figure 1: DS-11 Typical BCA Assay Standard Curve



Figure 2: DS-11 Typical Bradford Assay Standard Curve



## Results (continued)

Figure 3: DS-11 Typical Lowry Assay Standard Curve





Figure 4: DS-11 Typical 660 nm Assay Standard Curve



## Summary

The data presented in this technical bulletin demonstrate that common protein colorimetric assays can easily be performed using the DeNovix DS-11 microvolume spectrophotometer. The Standard Curve screen enables the user to measure between 2 to 8 standards and up to 5 replicates for each concentration. The EasyApp™ software allows the user to save routinely used concentration series or standard curve data sets.

The Run Screen (*not shown*) conveniently includes two graphs for sample measurements. One graph displays the sample absorbance spectrum and while the second graph displays where the sample falls on the standard curve.

The ease of use of the DS-11 makes this the ideal microvolume spectrophotometer for routine protein quantification using the four most common colorimetric assays.

### DeNovix, Inc.

3411 Silverside Road - Hanby Building  
Wilmington, DE 19810 USA

Phone: +1.302.442.6911  
Fax: +1.302.792.7407  
Email: info@denovix.com  
[www.denovix.com](http://www.denovix.com)

